

# Certificate of Quality

## Asbestos PCM & TEM Sampling Cassettes

**PCM Lot # 20100112**

OSHA Method ID-160 & NIOSH Method 7400  
(Phase Contrast Microscopy = PCM)  
or NIOSH Method 7402  
(Transmission Electron Microscopy = TEM)

- Clear with Acetone & Triacetin - Method 7400 (Quickfix)
  - o Filter materials Clear with smooth (non-grainy) background for accurate analysis.
  - o Time to Clear (seconds) <5 Seconds
- Background Fibers PCM 0.0 Fibers/100 Field and/or TEM NA
  - o Up to 4 fibers per 100 fields (ID-160) or
  - o Up to 5 fibers per 100 fields (NIOSH 7400 & NIOSH 7402)
- Conductive polymers (cowl) Per Resin Supplier
- Specify a consistent effective collection area of 385 +/- 5 mm<sup>2</sup>
  - o (21.998 mm <diameter> 22.283 mm collection area)
  - o Diameter AVG(100) = 22.16 = 385.86 mm<sup>2</sup>
- Leak Free Cassettes

	Open LPM	Inlet/Cowl/Outlet (AVG) LPM
o Rotary Vane	<u>35.00</u>	<u>19.96</u>
o Diaphragm	<u>27.30</u>	<u>16.21</u>
o Megalite	<u>30.84</u>	<u>14.66</u>

**PCM LOT # 20100112 - 0.8um Filter Membrane**

Approved: \_\_\_\_\_

*James R. Tucker*

Date: 01/12/2010

James R. Tucker, Ph.D.  
EMS COO & VP Technology



environmental monitoring  
s y s t e m s