

# Certificate of Quality

## Allergenco-D™ Disposable Air Sampler

**Airflow** - 15 liters per minute for 10 minutes

(Limited or Extended Times are recommended for highly contaminated or very clean environments, respectively)

**d50 cutoff** = 1.8 $\mu$ m

### Good Manufacturing Practices

This product was manufactured in a facility which meets or exceeds good manufacturing practice standards.

### Quality Assurance Lot Release Criteria

This manufacturing lot was tested and released by Quality Assurance to the following specifications:

#### Structural Integrity

Samples were found to be integral as measured by our analytical challenge test.

#### Collection Adhesive

100% of cassettes are visually examined to be >99.97% free of contaminating particles on the adhesive trace surface.

Validated by: *James R. Tucker*  
James R. Tucker, Ph.D.  
Vice President of Technology

Made in Charleston, SC USA



This certificate is issued as evidence of ems' commitment to provide only rigorously tested environmental products, equipment and supplies to our valued customers.

Environmental Monitoring Systems  
Toll Free: 1.800.293.3003 [www.emssales.net](http://www.emssales.net)